

For Brix				
Model	Scale	Range	Div.	Accuracy
DRB0-45	Brix	0-45%	0.1%	±0.2%
DRB0-45nD	Brix	0-45%	0.1%	±0.2%
	Refractive Index	1.3330-1.4098	0.0001	±0.0003
DRB28-65	Brix	28-65%	0.1%	±0.2%
DRB28-65nD	Brix	28-65%	0.1%	±0.2%
	Refractive Index	1.3770-1.4535	0.0001	±0.0003
DRB58-92	Brix	58-92%	0.1%	±0.2%
DRB58-92nD	Brix	58-92%	0.1%	±0.2%
	Refractive Index	1.4370-1.5090	0.0001	±0.0003
DRBS-300	Brix	0-35%	0.1%	±0.2%
	Salinity	0-28%	0.1%	±0.2%
	Refractive Index	1.3330-1.3900	0.0001	±0.0003

For Salinity				
Model	Scale	Range	Div.	Accuracy
DRS0-28	Salinity	0-28%	0.1%	±0.2%
DRS0-28nD	Salinity	0-28%	0.1%	±0.2%
	Refractive Index	1.3330-1.3900	0.0001	±0.0003
DRBS-300	Brix	0-35%	0.1%	±0.2%
	Salinity	0-28%	0.1%	±0.2%
	Refractive Index	1.3330-1.3900	0.0001	±0.0003

For Clinical Protein				
Model	Scale	Range	Div.	Accuracy
DRC-200	Urine Gravity	1.000-1.050	0.001	±0.001
	Serum Protein	0-12g/dl	0.1	±0.1
	Refractive Index	1.3330-1.3900	0.0001	±0.0003

For Battery/Antifreeze/Cleaning Fluid				
Model	Scale	Range	Div.	Accuracy
DRA-403C	Propylene Glycol	-95-0 °C	0.1 °C	±0.20 °C
	Ethylene Glycol	-70-0 °C	0.1 °C	±0.20 °C
	Battery	1.000-1.510sg	0.01sg	±0.01sg
	Cleaner	-72-0 °C	0.1 °C	±0.20 °C
DRA-403F	Propylene Glycol	-130-32 °F	0.1 °F	±0.5 °F

	Ethylene Glycol	-95-32 °F	0.1 °F	±0.5 °F
	Battery	1.000-1.510sg	0.01sg	±0.01sg
	Cleaner	-100-32 °F	0.1 °F	±0.5 °F

For °OE

Model	Scale	Range	Div.	Accuracy
DRB-400	MASS s/v	0-35%	0.1%	±0.2%
	Vol AP	0-22%	0.1%	±0.2%
	Oe	0-150	1.0	±1.0
	KMW	0-25	0.1	±0.2

For Bee Honey

Model	Scale	Range	Original Point	Div.	Accuracy
DRH-300	nD	1.4419-1.5232	1.4419(±0.0001)	0.0001	±0.0003
	°Be'	33.1-47.7	33.1	0.1	±0.2%
	WATER	38.4%-5.4%	38.4	0.1%	±0.2%