



## FiT239-Color Professional Breath Alcohol tester (Fuel Cell Sensor)



Display: TFT 2.4" TFT QVGA display with touch panel

Keyboard: 5 keys operation: "Up", "Down", "Left", "Right", "OK"

Power supply: Rechargeable/interchangeable battery (NOKIA standard)

Heat& warm up: Chamber & Sensor warm up automatically

Flow detect: Flow detect automatically

Operation environment: From -20°C to +70 °C, (-20 Degree C to +40 Degree C for optimum operation)

Data download: PC Download for IBM compatible PC

Memory: 1M Flash memories for Data log (option 1M max.), (3000 memories with ID & time stand)

Standard deviation +/-0.025 mg/L or +/-5 % of the measurement

Drift: typically < 0.8 % of the measurement value/month

Comply with S&E, CE, AS , NF, DOT (TBC)

Standard packing: Including metal case; charger cable; 2pcs mouthpieces

**Standard:**

Certified GB/T 21254-2007; CE  
Compile with S&E, AS , NF, DOT, OIML126