Huizhou Woroy Medical Technology Co., Ltd.

Jinshan Lake Industrial Zone, Huicheng District, Huizhou City, Guangdong,China

Mobile/WhatsApp/Wechat/Skype: +86 13068289263

Contact:Kevin E-mail：Sales02@woroy.com.cn

Product parameter information:(Digital Thermometer)

Number	Project	Content
1	Material	ABS, Stainless Steel
2	Display: Range	32.0-42.9℃.89.6-107.6°F
3	Max Error	0.1℃.0.2°F
4	Accuracy	±0.1℃ (35.5-42℃); ±0.2℃(<35.5℃).±0.2°F (95.5-107.6°F); ±0.6°F(<95.5°F)
5	Time response	60±10S(oral).100±20S(underarm)
6	Battery	DC1.5V TYPE-LR41

Special Note:

1. The above parameters are for reference, please refer to the instructions for details.

2. If you have better needs, we can customize various styles and technical parameters for you.